



Product Quality Control Report

Product Name: MIGRASEE Migrasome / Microscale Particle Observation Chip

Catalog No.: MGS-P101

Lot No.: 2024001

QC Item	Acceptance Criteria / Method	Result	Disposition
Product Inspection	The substrate is undamaged	Meets specification	Pass
	The four PDMS edges are parallel to the substrate; the end-margin difference along the long side is no more than 0.2 mm	Meets specification	Pass
	All four PDMS edges are fully bonded to the substrate	Meets specification	Pass
	Visual inspection confirms that no channel has top-to-bottom wall adhesion	Meets specification	Pass
	Five randomly selected channels are free of debris under microscopic inspection	Meets specification	Pass
Packaging Inspection	The product is completely wrapped, adequately cushioned, and secured inside the package	Meets specification	Pass
	The packaging is clean and undamaged	Meets specification	Pass
	Labeling is legible and clearly states the name, manufacturing date, shelf life, and precautions; the stated specification matches the contents	Meets specification	Pass
Conclusion	In accordance with the company's inspection criteria, the samples from this lot passed all tests		
QC Inspector Signature		Authorized Signature	
Date		Date	

Manufacture, processing, and reporting for this product conform to direct-volume standard MGSQCP101.

Representative samples were tested in accordance with Migrasome Therapeutics quality-control procedures and met the product performance specifications.

Catalog No.: MGS-P101

For products shipped before 2026-07-01

